

Oxymo Conserver

When your Oxymo conserver is cold or not connected to an oxygen cylinder, it will test itself. While it is testing, you may hear a fast clicking sound. This is normal and will stop once the conserver is warm, or when you connect it to your oxygen cylinder and turn the cylinder on to setting 4.



**Caution please
do not store the
conserver below
0° Celsius**

